

Certificate of Compliance

Certificate: 1240445

Master Contract: 212588

Project: 1240445

Date Issued: February 12, 2002

Issued to: **Hydropath UK Ltd.**
Unit F
Acorn Park Industrial Estate
Redfield Road
Nottingham NG7 2TR
UNITED KINGDOM

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by:

M. Brössoit
M. Brössoit, Eng.

PRODUCTS

CLASS 8721 04 - LABORATORY ELECTRICAL EQUIPMENT

CLASS 8721 84 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - CERTIFIED TO U.S. STANDARDS

HydroFLOW and AquaKLEAR physical water conditioner treatment system, rated 87-240V ac, 1A, 50/60Hz.

No.	Conditioner Model	System Name	Power Supply Model
1	C45, C60, C100, C120, C160, P60, P100 and P120	<i>HydroFLOW</i>	DEL 638 (module size: 125 x 80 x 140mm)
2	C45, C60, C100, C120, C160, P60, P100 and P120	<i>HydroFLOW</i> or <i>AquaKLEAR</i>	CPI (module size: 125 x 80 x 140mm)
3	P160	<i>AquaKLEAR</i>	CP2 (module size: 125 x 80 x 140mm)
4	Custom Models	<i>AquaKLEAR</i> or <i>HydroFLOW</i>	CP2 (module size: 125 x 80 x 140mm)
5	HS38B	<i>HydroFLOW</i>	AC-35-1210 (Pt T2808SA) (UK plug-in adapter)
6	HS38C	<i>HydroFLOW</i>	AC-35-1210 (Pt TS2807SA) Euro plug-in adapter)
7	HS38D	<i>HydroFLOW</i>	AC-35-1210 (Pt TS2806SA) (US plug-in adapter)
8	HS38A	<i>HydroFLOW</i>	AC-35-1210 (Pt T2693SA) (UK/Euro in-line adapter)
9	HS38E	<i>HydroFLOW</i>	AC-35-1210 (Pt TS2809SA) (US in-line adapter)

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.

Certificate: 1240445

Master Contract: 212588

Project: 1240445

Date: February 12, 2002

APPLICABLE REQUIREMENTS

- CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)
- UL Std No. 3101.1 - Electrical Equipment for Laboratory Use; Part 1: General Requirements
- IEC 61010-1 (1990) - Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use - Part 1: General Requirements
Amendment No. 1 (1992)
Amendment No. 2 (1995)



Supplement to Certificate of Compliance

Certificate: 1240445

Master Contract: 212588

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
1240445	Feb. 12, 2002	Original Certification.